

CERTIFICATE OF ANALYSIS

General information

HRD-P13-T (Tumor)			
CBP90023-13T			
CBFA24072323X			
Low 25=5(LOH)+10(TAI)+10(LST) with Panel+SNP method;			
Low 19=6(LOH)+10(TAI)+3(LST) with WGS method;			
Genomic DNA			
Tris-EDTA			
2~8°C			
2024/07/23			
36 months from the date of manufacture			

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	V
Concentration	40.00 ng/uL±10%	43.00 ng/uL by Qubit 3.0 detection	V
OD_{260}/OD_{280}	1.7~2.1	1.84 by Nanodrop detection	
DNA electrophoresis	No Degradation	10000 bp — 8000 tp — 80000 bp — 800000 bp — 80000 bp —	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 彭美楠 Date:2025/01/07