


CERTIFICATE OF ANALYSIS

General information

| | |
|---------------------|---|
| Name | Package-RefHRD Cocktail Reference Standard CBP90023-9-N(normal) |
| Cat. No. | CBP90023-9-N |
| Lot No. | CBFA23070404 |
| Format | Genomic DNA |
| Buffer | Tris-EDTA |
| Storage Conditions | -25°C~ -15°C |
| Date of Manufacture | 2023/07/04 |
| Expiry | 36 months from the date of manufacture |


Quality Control

| Test Items | Standard | Results | Passed |
|--------------------------------------|----------------------|---|--------|
| Quantity | 1.00 µg/vial*5 vials | 1.00 µg/vial*5 vials | √ |
| Concentration | 40.00 ng/uL±10% | 44.00 ng/uL by Qubit 3.0 detection | √ |
| OD ₂₆₀ /OD ₂₈₀ | 1.7~2.1 | 1.83 by Nanodrop detection | √ |
| DNA electrophoresis | No Degradation |  <p>M: Marker 1: Package-RefHRD Cocktail Reference Standard CBP90023-9-N(normal)</p> | √ |

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:2025/03/06

 Date:2025/03/06