


CERTIFICATE OF ANALYSIS

General information

Name	HRD-P1-T (tumor)
Cat. No.	CBP90023-1
Lot No.	CGFA23050608
HRD Score	High 95=28(LOH)+34(TAI)+33(LST) with Panel+SNP method; High 70= 17(LOH)+27(TAI)+26(LST) with WGS method;
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	2~8°C
Date of Manufacture	2023/05/06
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	42.80 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.81 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: HRD-P1-T (tumor)</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 彭亚楠 Date:2025/05/13

刘明 Date:2025/05/13